

SCHEDULE XIX

Minimum norms and standard for Atyayika (Emergency) Outpatient department and Intensive Care Unit

Serial number	Standards or requirement	Intake capacity wise minimum requirements				Availability Quantity	Deficiency
		60	100	150	200		
Atyayika (Emergency) outpatient department with Observation Beds							
1.	Motorized bed four section with mattress	2	2	4	4	2	
2.	Wheelchair	1	1	2	2	1	
3.	Bedsider locker	2	2	4	4	2	
4.	Overbed table	2	2	4	4	2	
5.	Intra-venous stand with stainless steel rod and castor base	2	2	4	4	2	
6.	Footstep double	2	2	4	4	2	
7.	Multiparamonitor	2	2	4	4	2	
8.	Intensive care unit ventilator	2	2	4	4	2	
9.	Portable monitor	1	1	3	3	1	
10.	Portable ventilator	1	1	2	2	1	
11.	Portable X-ray	1	1	1	1	1	
12.	Patient stretcher	1	1	2	2	1	
13.	Dressing trolley	1	1	2	2	1	
14.	Drug trolley or medicine cart	1	1	2	2	1	
15.	Electrocardiogram machine trolley	1	1	2	2	1	
16.	X-Ray view box	1	1	1	1	1	
17.	Suction machine	1	1	2	2	1	
18.	Suturing set	2	2	3	5	2	
19.	Agnikarmakrit	1	1	1	1	1	
20.	Dhumayantra	1	1	2	2	1	
21.	Naadi recording equipment	1	1	2	2	1	
22.	Sthanikswedanayantra	1	1	2	2	1	
23.	Common diagnostic tools such as- Non-mercurial sphygmomanometer, stethoscope, torch, clinical thermometer (non-contact), tongue depressor weight and height measuring stand, measuring tape	As required				As required	
24.	Emergency medicine	As required				As required	


Principal
 Shamrao Patil Institute of Ayurved
 Research Centre And Hospital Paniv,
 Tal. Malshiras, Dist. Solapur-413113